

**Tydd St Mary Parish Council**

**Procedure following Discharge of Defibrillator**

If possible, request a replacement set of pads from paramedics.

**A**

If pads obtained:

Replace new pads tray in defibrillator and run through check procedure.

If unit functioning and rescue ready replace in cabinet (if not follow B instructions)

Contact Clerk for replacement rescue kit.

**B**

If pads not obtained:

Put Sign in defibrillator cabinet informing out of use and location of nearest alternative.

Put defibrillator in a safe place (NOT back in cabinet)

Contact Clerk and inform of defibrillator discharge = (tyddstmarypc@gmail.com)

Clerk will inform EMAS & the Circuit to de-activate CPAD until informed site is rescue ready.

Clerk will order new pads.

(then follow A instructions)